

Metals by EPA 200.7 (Waste Water)

Parameter	MRL (mg/l)
Silver	0.005
Aluminum	0.05
Arsenic	0.015
Boron	0.02
Barium	0.01
Beryllium	0.001
Calcium	0.1
Cadmium	0.001
Cobalt	0.005
Chromium	0.005
Copper	0.005
Iron	0.1
Potassium	0.5
Magnesium	0.1
Manganese	0.005
Molybdenum	0.005
Sodium	0.5
Nickel	0.005
Lead	0.005
Antimony	0.015
Selenium	0.015
Thallium	0.015
Vanadium	0.005
Zinc	0.01
Tin	0.1
Silica	0.1
Titanium	0.01

Metals200.7_041108

* For samples requiring more sensitive Reporting Limits (RL), it is recommended to use EPA Method 200.8